

Report

103-76-4.qgd
D:\ding_ke\pql\103-76-4.qgd
Admin 2015-9-21 13:44:57

Method

[Comment]

===== Analytical Line 1 =====

[GC-2010]

Column Oven Temp. :60.0 °C
Injection Temp. :220.00 °C
Injection Mode :Split
Flow Control Mode :Pressure
Pressure :93.1 kPa
Total Flow :79.5 mL/min
Column Flow :1.50 mL/min
Linear Velocity :44.7 cm/sec
Purge Flow :3.0 mL/min

Oven Temp. Program


Rate	Temperature(°C)	Hold Time(min)
-	60.0	0.00
10.00	100.0	4.00
10.00	200.0	20.00
10.00	250.0	10.00

\$EndIf\$

[GCMS-QP2010 Plus]

IonSourceTemp :250.00 °C
Interface Temp. :250.00 °C
Solvent Cut Time :1.00 min

Peak#	R.Time	I.Time	Peak Report TIC		Area%	Height	
			F.Time	Area		Height	Height%
1	7.939	7.683	8.375	101380074	100.00	8521303	100.00
				101380074	100.00	8521303	100.00



Spectrum

Line#:1 R.Time:7.9(Scan#:714)

MassPeaks:57

RawMode:Averaged 7.9-8.0(713-715) BasePeak:99(1736445)

BG Mode:Calc. from Peak Group 1 - Event 1

